



# Validation of Illumina TruSeq® Amplicon Cancer Panel with Concordance Testing Using Ion AmpliSeq Cancer Panel and Other Methods

Peng Fang, Zhenyu Yan, Weihua Liu, Agus Darwanto, Kim Pelak, Kim Anoe, Cindy Spittle, Sabita Sankar, Chad Galderisi, and Jin Li, MolecularMD Corp., Cambridge, MA and Portland, OR

## Introduction

The Illumina TruSeq® Amplicon Cancer Panel (TSACP), a highly multiplexed targeted resequencing assay for use on the Illumina MiSeq platform, is designed for detecting the hotspot mutations in 212 Regions of Interest (ROI) from 48 cancer-related genes (Table 1). We used well characterized cancer cell lines harboring clinically relevant variants as positive controls and HapMap samples NA12878 and NA19240 as wild type control samples. A total of 41 FFPE patient specimens representing a variety of cancer types were analyzed in a blinded fashion to evaluate the analytical sensitivity and specificity of the assay. With no gold standard available as a reference method to detect mutations with comparable sensitivity, concordance testing was performed using the Ion AmpliSeq Cancer Panel. Variants detected by both panels were considered as true positives. Variants that were either not covered by the Ion Panel or not concordantly detected were confirmed by a third method, either Sanger sequencing (variants with frequencies > 10%) or a MolecularMD custom Ion TargetSeq Assay (variants with frequencies <10%).

Table 1. The 48 genes targeted by Illumina TSACP assay

ABL1	BRAF	EGFR	FGFR2	GNAS	JAK3	MLH1	PDGFRA	RET	STK11
AKT1	CDH1	ERBB2	FGFR3	HNF1A	KDR	MPL	PK3CA	SMAD4	TP53
ALK	CDKN2A	ERBB4	FLT3	HRAS	KIT	NOTCH1	PTEN	SMARCB1	VHL
APC	CSF1R	FBXW7	GNA11	IDH1	KRAS	NPM1	PTPN11	SMO	
ATM	CTNNB1	FGFR1	GNAQ	JAK2	MET	NRAS	RB1	SRC	

## Materials and Methods

The DNA samples were quantified with Qubit 2.0 (Life Technologies), and the quality of the FFPE DNA was assessed with a qPCR based, Infinium HD FFPE QC Kit (Illumina). 150 ng genomic DNA or 250 ng qualified FFPE DNA was hybridized to the pre-mixed oligonucleotide probes that flank the 212 ROI. An extension reaction extends across the ROI, followed by ligation to unite probe pairs, creating new template strands. The templates generated from extension-ligation were PCR-amplified with the primers adding the index sequences for sample multiplexing as well as the common adapters required for the cluster generation. The indexed library products were purified with AMPure XP (Beckman-Coulter), normalized to equal concentration, pooled and loaded on MiSeq for automated cluster amplification and sequencing. Data were analyzed using MiSeq Reporter software and our proprietary analysis pipeline. In addition to reporting hotspot mutations, we also report "Critical Variants" such as non-synonymous coding mutations and splicing site mutations that fall within the ROI.

## Results

**ROI Coverage:** In the seven sequencing runs (16 samples per flow cell) performed during our validation study, the mean sequencing coverage of each run ranged from 1607 to 4029 (Table 2). Nine ROI, or 4.4% of total ROI, were consistently sequenced with less than 0.2x normalized coverage (Table 2 and Table 3, highlighted green), whereas between 4 and 23 ROI had less than 500x read depth, or 2.2% to 12.6% of total ROI, depending on the mean coverage of each run (Table 3, highlighted orange).


Table 2. Summary of ROI coverage

Run#	Average coverage per sample	DNA loading on MiSeq (cluster density, k/mm <sup>2</sup> )	% ROI >500 depth	% ROI <0.2x coverage
Run#3	1607	443	12.6%	
Run#4	1683	417	12.1%	
Run#1	1750	398	11.5%	
Run#6	2480	629	4.9%	4.4%
Run#5	2564	631	4.4%	
Run#2	4007	996	2.2%	
Run#7	4029	1095	2.2%	

Table 3. ROI covered at less than 500x read depth

Region of Interest	Gene	Normalized coverage	Average sequencing depth per ROI						
			Run#3	Run#4	Run#1	Run#6	Run#5	Run#2	Run#7
Average coverage per sample			1607	1683	1750	2480	2564	4007	4029
9:21971149-21971282	CDKND1	0.01	10	10	10	15	15	24	24
4:1808318-1808442	FGFR3_8	0.10	154	161	168	238	246	384	386
20:36031762-36031895	SRC	0.11	184	193	201	284	294	459	462
19:3118827-3118963	GNA11_5	0.12	189	198	206	292	302	471	474
3:10183795-10183926	VHL1_2	0.14	228	239	249	352	364	569	573
7:128851591-128851708	SMO5	0.15	244	256	266	377	390	609	612
4:1803564-1803685	FGFR3_1_2	0.16	257	269	279	396	409	640	643
7:128846374-128846509	SMO3	0.18	284	298	309	439	453	708	712
9:13378329-13378465	ABL1_1	0.20	321	337	350	496	513	801	806
13:48955534-48955657	RB1_4	0.23	369	386	401	569	588	919	924
9:139397767-139397901	NOTCH1_1	0.24	387	405	422	597	617	965	971
19:3114847-3115111	GNA11_4	0.25	401	420	436	618	639	999	1004
11:534279-534410	HRAS6	0.25	409	428	446	631	653	1020	1026
4:1807859-1807999	FGFR3_5_7	0.26	414	433	451	639	660	1032	1038
4:1806119-1806250	FGFR3_3_4	0.26	422	442	460	651	673	1052	1058
4:5514456-55144676	PDGFRA6	0.27	428	449	466	661	683	1068	1074
9:133747504-133747641	ABL1_9	0.27	429	449	467	662	684	1069	1075
17:3780898-37881124	ERBB2_4	0.27	432	452	470	666	688	1076	1082
10:43613834-43613971	RET10	0.27	442	463	481	682	705	1101	1108
13:28610124-28610239	FLT3_22	0.28	452	473	492	697	720	1126	1132
19:1220486-1220614	STK11_4	0.28	457	478	497	705	729	1139	1145
7:128850336-128850472	SMO4	0.30	476	498	518	734	759	1187	1193
1:115256527-115256653	NRAS1_7	0.31	492	515	535	759	784	1226	1233
11:108218089-108218220	ATM17	0.33	534	559	581	823	851	1330	1338

Figure 1. FFPE sample QC summary



**Mutations Identified:** Among 38 samples (37 FFPE and 1 mixed cell line DNA), with 5% LOD cut-off, 124 unique critical mutations (mutations that would affect the protein sequence) were identified. These included 107 single base substitutions (SBS, Figure 2), 14 unique single- or multi-base deletions (Table 4), and 3 single- or two- base insertions (Table 5). The mutation spectrum in each tumor type is shown in Figure 3.

Figure 2. The number of SBS mutations identified




Table 4. Mutations caused by deletions


Sample Name	Tumor Type	Chromosome coordinate	Gene name	Mutated CDS	Mutated Amino Acids	Freq (%)	Read Depth
MMD-2-4-002	GIST	4:55152092-55152103	PDGFRA	c.2524_2535delGGA/CATCATGCAT	p.Leu843_Asp846del	25.9	8125
MMD-2-4-003	GIST	4:55593589-55593594	KIT	c.1655_1660delTTG/TATG	p.Met552_Glu554delinsyls	60.0	1513
MMD-2-4-004	GIST	4:55593640-55593660	KIT	c.1706_1726delTTTAA/CATAGACCCCAACACAAAC	p.Tyr570_Leu576del	37.0	5682
MMD-2-4-023	colon	4:55595655	KIT	c.2045delT	p.Leu682X	29.4	2811
MMD-2-4-025	colon	5:112175923	APC	c.4632delA	p.Glu1544A>GpfsQ21	86.4	708
Mix-6can	Cell line	7:55242465-55242479	EGFR	c.2235_2249delGGAAATTAAGAGAGAGC	p.Glu746_Ala750del	10.5	3060
MMD183	colon	9:133748412	ABL1	c.11073delA	p.Asn358T>TrfsX14	9.0	4503
MMD183	colon	10:89711914-89711916	PTEN	c.532_534delTAT	p.Tyr178del	23.8	1032
MMD151	endometrium	10:89717758	PTEN	c.783delG	p.Asn262>TrfsX4	97.8	2016
MMD-2-3-063	ovarian	10:89720771	PTEN	c.922delC	p.Arg308Val>fsX9	42.0	7811
MMD-2-3-052	endometrium	10:89720804-89720807	PTEN	c.955_958delACTT	p.Thr319X	52.5	4647
MMD-2-3-053	endometrium	10:89720804-89720807	PTEN	c.955_958delACTT	p.Thr319X	46.6	4522
MMD-2-4-023	colon	18:48581289	SMAD4	c.593delC	p.Pro198Gln>fsX4	26.0	4686
Mix-6can	Cell line	19:1207081	STK11	c.169delG	p.Glu57Lys>fsX7	11.2	206
Mix-6can	Cell line	19:1221319	STK11	c.842delC	p.Pro281Arg>fsX6	30.2	2032

Table 5. Mutations caused by insertions

Sample Name	Tumor Type	Chromosome coordinate	Gene name	Mutated CDS	Mutated Amino Acids	Freq (%)	Read Depth
MMD-2-4-020	colon	5:112175480-112175481	APC	c.4189dupG	p.Glu1397Gly>fsX12	31.6	1372
MMD183	colon	5:112175685-112175686	APC	c.4393_4394dupAG	p.Ser1465Arg>fsX9	26.2	6167
MMD-8-2013	endometrium	10:89720803-89720804	PTEN	c.954dupT	p.Trp319Tyr>fsX6	61.3	5547


**Analytical Reproducibility:** The analytical reproducibility of the TSACP assay was evaluated with a sample created by mixing DNA from 6 cancer cell lines (sample ID = Mix-6can) and DNA extracted from 13 FFPE tissue samples. With a 5% allele frequency cut-off, 100 of 102 expected mutations were reproducibly detected within or between runs (Figure 4). The two missed mutations were still correctly called at less than 5% allele frequency in those two samples, suggesting a variation of minor allele frequency determined by TSACP assay. The CV for the minor allele frequency of these mutations ranges from 0-38%, with the majority (88%) below 20%.

Figure 4. Analytical reproducibility determined with DNA from 13 FFPE specimens and 1 cell line



**Confirmation of Critical Mutations Identified:** For the 139 critical mutations identified in FFPE samples, 104 (75%) were cross-confirmed by the Ion Torrent AmpliSeq Cancer Panel Assay (Figure 5), while 32 were not covered by the ROI of the Ion assay. Of the 35 mutations either not covered by, or not concordant with the Ion assay, 3 mutations were not confirmed due to insufficient DNA available; 23 were analyzed by Sanger sequencing (all 23 were true positives), and 9 were analyzed by a MolecularMD custom Ion TargetSeq assay (5 of the 9 were true positives, Table 6). These same variant positions are also shown for the wild-type control NA12878. Of the 4 mutations called by the TSACP assay that were found to be false positives, 3 were located in the GNA11 gene which had less than 200x read depth (Table 6, highlighted).

Figure 5. Mutation confirmation summary



**Sensitivity and Specificity of the TSACP assay:** The results shown above indicate that the sensitivity and specificity of the TSACP assay are 99% and 97%, respectively (Table 7). To further enhance the specificity of the assay, a stringent variant filter could be applied, requiring that the mutated nucleotide has to be covered by at least 25x read depth in order to be reported. This filter removes 3 FP mutations and 1 TP mutation. This stringent coverage filter reduces the FP rate from 3% to 1%, with 2 FN identified. Consequently, with the filter the sensitivity and specificity are 98% and 99% respectively (Table 7).

Table 7. Sensitivity and specificity of the TSACP assay

	No Filter	Filter
Number of mutations reported without confirmatory test	136	132
True Positive (TP)	132	131
False Positive (FP)	4	1
Number of False Negative (FN) mutations missed	1	2
<b>Sensitivity (TP/(TP+FN))</b>	<b>99%</b> (132/(132+1))	<b>98%</b> (131/(131+2))
<b>Specificity (TP/(TP+FP))</b>	<b>97%</b> (132/(132+4))	<b>99%</b> (131/(131+1))

Table 6. MolecularMD Custom Ion TargetSeq analysis for critical mutations below Sanger detection limit

Sample name	Chromosome coordinate	Gene name (HGNC)	Mutated CDS	Mutated Amino Acids	TSACP		MMD-Ion TargetSeq		Result
					Allele Freq (%)	Read Depth	Allele Freq (%)	Read Depth	
MMD-2-2-025	3:10191533	VHL	c.526A>T	p.Arg176Trp(176R>W)	12.4	2714	13.4	47222	TP
MMD-2-3-051	4:55144636	PDGFRA	c.2110G>T	p.Glu1704G(1704E>*)	9.1	561	12.8	7934	TP
MMD-2-4-003	10:43613892	RET	c.2356G>T	p.H8786Tyr(786H>Y)	5.4	594	0.0	45443	FP
MMD-2-3-051	13:49027202	RB1	c.1769G>C	p.Cys590Ser(590C>S)	10.2	763	10.1	28068	TP
MMD-2-3-051	18:48575206	SMAD4	c.400G>A	p.Glu134Lys(134E>K)	10.5	2134	9.2	21036	TP
MMD-2-3-052	19:3114952	GNA11	c.487G>A	p.Asp163Asn(163D>N)	8.3	168	0.0	5681	FP
MMD-2-5-021	19:3114962	GNA11	c.497G>A	p.Arg166His(166R>H)	5.8	120	0.0	1432	FP
MMD-2-4-003	19:3114967	GNA11	c.502G>A	p.Ala168Thr(168A>T)	5.6	195	0.0	2644	FP
MMD-2-3-051	19:3115036	GNA11	c.571G>A	p.Glu191Lys(191E>K)	13.0	1073	16.6	21984	TP
NA12878 (wild-type control)	3:10191533	VHL	c.526A>T	p.Arg176Trp(176R>W)	0.0	0	0.0	79589	
	4:55144636	PDGFRA	c.2110G>T	p.Glu1704G(1704E>*)	0.0	0	0.0	23737	
	10:43613892	RET	c.2356G>T	p.H8786Tyr(786H>Y)	0.0	0	0.0	22460	
	13:49027202	RB1	c.1769G>C	p.Cys590Ser(590C>S)	0.0	0	0.0	28524	
	18:48575206	SMAD4	c.400G>A	p.Glu134Lys(134E>K)	0.0	0	0.0	13488	
	19:3114952	GNA11	c.487G>A	p.Asp163Asn(163D>N)	0.0	0	0.0	2548	
	19:3114962	GNA11	c.497G>A	p.Arg166His(166R>H)	0.0	0	0.0	6428	
	19:3114967	GNA11	c.502G>A	p.Ala168Thr(168A>T)	0.0	0	0.0	2548	
	19:3115036	GNA11	c.571G>A	p.Glu191Lys(191E>K)	0.0	0	0.0	16327	

## Conclusions

- Our validation study demonstrated that the Illumina TruSeq® Amplicon Cancer Panel combined with MolecularMD's proprietary analysis pipeline provides a highly specific and sensitive test suitable for screening patient FFPE tumor specimens for a spectrum of clinically relevant somatic mutations with analytical sensitivity of 5% minor allele frequency.
- With more than 95% ROI sequenced at least 0.2x normalized coverage, the analytical reproducibility was above 98%.
- In addition to reporting hotspot mutations, we also report other variants that would impact protein sequence. We report all mutations in our ROI, except for those characterized as synonymous or intronic.
- The diagnostic sensitivity and specificity for FFPE samples, under these validation conditions, are 98% and 99% respectively when a stringent coverage filter is applied during data analysis.
- In this study, we developed a unique cross-validation strategy using two different NGS platforms. We also established a custom Ion TargetSeq® assay for confirming variants at frequencies below 10%. In total, 98% of all the mutations identified in FFPE samples were cross-examined by other platforms, i.e., the Ion Torrent AmpliSeq Cancer Panel, the custom Ion TargetSeq, or Sanger sequencing.

For Further Information please contact [info@molecularmd.com](mailto:info@molecularmd.com) or visit [www.molecularmd.com](http://www.molecularmd.com).

