



Mutation Profiling of cfTNA using Oncomine Pan-Cancer Cell-Free Assay

Xiaodong Wang, Zhenyu Yan, Weihua Liu, Scott Glynn, Jennifer Biroschak, Chad Galderisi, Cindy Spittle and Jin Li - MolecularMD Corp., Portland OR and Cambridge MA

Introduction

Liquid biopsy using cfDNA isolated from the plasma fraction of whole blood is less invasive for patients compared to tumor tissue biopsies. Targeted NGS analysis of cfDNA from cancer patients can be valuable in clinical trials for determining the presence of somatic mutations especially when a tumor tissue biopsy is not available. Cell-free RNA (cfRNA) analysis can also be used to detect the presence of gene fusions. The ability to analyze both cfRNA and cfDNA in a single assay provides an opportunity for comprehensive mutation profiling of cancer patients. The Oncomine Pan-Cancer Cell-Free Assay is designed to detect multiple targets in tumor-derived cfTNA (DNA and RNA) isolated from the plasma fraction of whole blood. The assay covers 52 genes, targeting hotspot single nucleotide variants (SNVs), short indels, gene fusions, copy number variations (CNVs), MET exon 14 skipping, and tumor suppressor genes (Table 1). Here we performed a proof of concept study to evaluate this TNA-based NGS assay using cfTNA controls and matched plasma and FFPE samples from late stage cancer patients.

Methods

cfTNA was isolated from the plasma using the MagMAX Cell-Free Total Nucleic Acid Isolation Kit. DNA and RNA from matched FFPE tumor tissue were isolated using the RecoverAll Total Nucleic Acid Isolation Kit. Control samples with 0.1% SNV/indel (positive control 1) and the CNV/Fusion/MET exon skipping (positive control 2) as well as the negative control were provided by the Thermo Fisher. The Structural Multiplex cfDNA Reference Standard and Multiplex I cfDNA Reference Standard Set were purchased from HorizonDx. 20 ng cfDNA or cfTNA was amplified using the Oncomine Pan-Cancer Cell-Free Assay following standard library construction to produce barcoded libraries. Four libraries were multiplexed for templating using standard protocols on the Ion S5 system with the Ion 540 chip. Sequencing data were analyzed with Torrent Suite software version 5.6 and Ion Reporter version 5.6 using Oncomine TagSeq Liquid Biopsy workflow. DNA and RNA from FFPE samples were analyzed using the Oncomine Comprehensive Assay v3 (OCAv3). Plasma and FFPE samples with discordant results between Pan-Cancer Cell-Free Assay and OCAv3 were further confirmed by the Oncomine Breast cfDNA Assay v2 and ddPCR assay by targeting PIK3CA and BRAF variants, respectively.

Results

Assay Performance and Limit of Detection

To validate the assay performance, we first analyzed 2 positive controls from Thermo Fisher, which include 87 SNVs and small indels in 25 genes at ~ 0.1% MAF (positive control 1), CNVs from 2 genes and fusions from 3 genes (positive control 2) (Table 1). The assay detected 76/87 (87%) SNVs from positive control 1. Two of the expected copy number gains were detected (MET and EGFR). All 3 expected gene fusions, i.e. EML4-ALK, SLC34A2-ROS1, CCDC6-RET as well as MET exon 14 skipping were also detected. No mutation was detected in the negative control.

We also challenged the assay using HorizonDx's Structural Multiplex cfDNA Reference Standard. The assay detected the expected copy number gain in MET and MYC as well as the SNVs in low GC and high GC regions. Interestingly, the assay also detected the expected fusion in SLC34A2-ROS1, CCDC6-RET from RNA portion of the assay although this standard is called cfDNA. This is likely caused by the inclusion of RNA in this commercial DNA standard (Table 2). In addition, 8/8 mutations in the HDx cfDNA reference standard were detected by the Oncomine Pan-Cancer Cell-Free Assay at 5%, 1% and 0.1%, except the 0.1% NRAS A59T (Table 3). In most cases, the limit of detection (LoD) was ~0.1% MAF as expected (Figure 1) when 20 ng cfDNA was used.

Table 2. Variant Detection in the HorizonDx Structural Multiplex cfDNA Reference Standard

Variant type	Gene	Variant	Expected	MAF (SNV/Indel) Detected or CNV ratio (inferred copy number)	Mutant Molecular Coverage
SNV	EGFR	G719S	5.3%	5.72%	131
SNV	BRAF	V600E	18.2%	16.61%	202
SNV	PIK3CA	H1047R	16.7%	16.81%	419
SNV High GC	GNA11	Q209L	5.6%	5.70%	172
SNV High GC	AKT1	E17K	5.0%	4.77%	116
SNV low GC	PIK3CA	E545K	5.6%	5.39%	136
SNV low GC	KRAS	G13D	5.6%	5.36%	160
Long insertion	EGFR	V769_D770insASV	5.6%	4.70%	112
Long deletion	EGFR	E746_A750del	5.3%	6.01%	123
CNV	MET	AMP	4.5 copies	1.67 (3.34 copies)	NA
CNV	MYC	AMP	9.5 copies	3.19 (6.38 copies)	NA
Fusion	ROS1	SLC34A2(4)-ROS1(32)	5.6%	NA	36
Fusion	RET	CCDC6(1)-RET(12)	5.0%	NA	16

Table 3. SNV Limit of Detection based on HorizonDx's cfDNA Reference Standards

Gene	Variant	Wild Type		5%		1%		0.1%	
		Expected	Detected AF	Expected MAF	Detected MAF (Mutant Molecular Coverage)	Expected MAF	Detected MAF (Mutant Molecular Coverage)	Expected MAF	Detected MAF (Mutant Molecular Coverage)
EGFR	L858R	0.00%	ND	5.00%	3.33% (85)	1.00%	0.70% (15)	0.10%	0.29% (7)
EGFR	ΔE746-A750	0.00%	ND	5.00%	5.65% (143)	1.00%	1.43% (30)	0.10%	0.12% (3)
EGFR	T790M	0.00%	ND	5.00%	3.61% (127)	1.00%	0.87% (28)	0.10%	0.12% (4)
EGFR	V769-D770 insASV	0.00%	ND	5.00%	4.38% (138)	1.00%	0.57% (15)	0.10%	0.11% (3)
KRAS	G12D	0.00%	ND	6.30%	7.17% (202)	1.30%	0.91% (22)	0.13%	0.19% (5)
NRAS	Q61K	0.00%	ND	6.30%	6.97% (137)	1.30%	1.66% (29)	0.13%	ND
NRAS	A59T	0.00%	NC	6.30%	NC	1.30%	NC	0.13%	NC
PIK3CA	E545K	0.00%	ND	6.30%	6.94% (185)	1.30%	1.38% (31)	0.13%	0.09% (2)

ND, Not detected; NC, Not covered by the assay.




Figure 1. Expected Assay Limit of Detection (From Thermo Fisher)

Table 1. Oncomine Pan-Cancer Cell Free Assay
NOTE: Gene variants in bold are included in positive control 1 and 2

Hotspot											Tumor suppressor			CNV		Fusion				
Genes	Variants	Mut Freq % (mut mol count)	Genes	Variants	Mut Freq % (mut mol count)	Genes	Variants	Mut Freq % (mut mol count)	Genes	Variants	Mut Freq % (mut mol count)	Genes	Variants	Mut Freq % (mut mol count)	Genes	Detection Copies	Genes	Detection (mut mol count)		
AKT1	E17K	0.137 (6)	ESR1			IDH2	R172K	0.175 (11)			S566_E571delinsR	0.308 (11)			R805*	0.082	CCND1	ALK	ELM4(6)-ALK(20) (12)	
ALK	F1174L	0.136 (8)	FGFR1				V530I	0.217 (8)	PDGFRA	N659K	0.065 (3)				R876*	0.154 (8)	CCND2	BRAF		
AR				N549K	0.119 (6)		K642E	0.212 (11)		D842V	0.087 (5)				S1234fs	0.130 (5)	CCND3	ERG		
ARAF				C382R	0.178 (8)	KIT	V654A	0.212 (11)		N848K	0.105 (6)				Q1291*	0.113 (4)	CDK4	ETV1		
	V600E	0.206 (6)	FGFR2	Y375C	0.178 (8)		H697Y	0.162 (5)		P539R	0.206 (11)				Q1294*	0.113 (4)	CDK6	FGFR1		
	L597R	0.206 (6)		G305R	0.255 (10)		G12A	2.082	PIK3CA	E542K	0.206 (11)				E1353*	0.081 (3)	EGFR	3.1	FGFR2	
	D594G	0.206 (6)		S252W	0.230 (11)	KRAS	G12D	0.088 (5)		E545K	0.206 (11)				Q1378*	0.081 (3)	ERBB2		FGFR3	
	N581S	0.206 (6)		S249C	0.268 (11)		K117N	0.088 (5)		H1047R	0.107 (5)				Q1406*	0.100 (5)	FGFR1		MET	MET(13)-MET(15) (501)
	G464V	0.114 (6)		G370C	0.132 (5)		Q61H	0.165 (8)	RAF1						R505C	0.075	FGFR2		NTRK1	
CHEK2				G380R	0.132 (5)	MAP2K1				C618R	0.071 (4)	FBXW7	WS26R	0.075 (4)	FGFR3	2.74	NTRK3			
	T41A	0.241 (11)	FGFR3	F384L	0.132 (5)	MAP2K2				C620R	0.071 (4)		R479Q	0.066 (4)	MET	2.5	RET		CCDC6(1)-RET(12) (35)	
CTNNB1	S45F	0.241 (11)		A391E	0.132 (5)		Y1253D	0.178 (7)	RET	E768D	0.150 (8)	PTEN			MYC				SLC34A2(4)-ROS1(32) (343)	
DDR2				G697C	0.273 (12)	MET	M1268T	0.178 (7)		A883F	0.193 (7)						ROS1		SLC34A2(4)-ROS1(34) (24)	
	R108K	0.157 (8)	FGFR4			MTOR				M918T	0.366 (14)				R248L	1.280				
	A289V	0.151 (8)	FLT3	D835Y	0.217 (13)		Q61R	0.255 (11)	ROS1						P278L	0.067 (2)				
	V292L	0.151 (8)	GNA11	Q209L	0.194 (12)	NRAS	G12D	0.185 (10)	SF3B1						R273H	0.067 (2)				
	p.L755S	0.150 (7)	GNAQ			NTRK1			SMAD4						G245S	0.083 (3)				
	p.D769Y	0.150 (7)	GNAS	R201C	0.124 (7)	NTRK3				L412F	0.204 (11)				Y220C	0.112 (6)				
	p.Glu770_Ala771insAVVM	0.129 (7)	HRAS	Q61R	0.085 (4)				SMO	W535L	0.113 (5)				V216M	0.112 (6)				
	p.V842I	0.201 (10)		G12V	0.142 (8)															
ERBB3			IDH1	R132H	0.230 (15)															

Assay Reproducibility

The HorizonDx 1% cfDNA Reference Standard was chosen to perform a repeat analysis to study the assay reproducibility (Table 4). The repeat analysis showed good assay reproducibility with similar MAF in each run.

Table 4. Assay Reproducibility

Gene	Variant	Expected	Detected MAF (Mutant Molecular Coverage)	
			A	B
EGFR	L858R	1.00%	0.7% (15)	0.2% (5)
EGFR	ΔE746 - A750	1.00%	1.4% (30)	0.8% (17)
EGFR	T790M	1.00%	0.9% (28)	0.5% (15)
EGFR	V769 - D770insASV	1.00%	0.6% (15)	0.6% (15)
KRAS	G12D	1.30%	0.9% (22)	1.4% (34)
NRAS	Q61K	1.30%	1.7% (29)	1.5% (25)
NRAS	A59T	1.30%	NC	NC
PIK3CA	E545K	1.30%	1.4% (31)	1.6% (34)

NC, Not covered.

Profiling of Plasma Samples from Cancer Patients

cfTNA extracted from seven patient plasma samples was analyzed by the Oncomine Pan-Cancer Cell-Free Assay. The cfTNA input was based on yield and varied from 8-28ng. All variants detected at >0.1% were reported for the purposes of this pilot study. Plasma cfTNA results were compared to matched FFPE tissue results and/or to cfTNA results using the Oncomine Breast cfDNA Assay or ddPCR assay. No variants were detected in plasma cfTNA using either NGS assay for samples 5 and 6. Three variants that were detected in the plasma cfTNA at >1% MAF were also detected in FFPE tissue (EGFR, BRAF V600K, NRAS Q61R, TP53 R280G). Three variants (italic) that were detected in the plasma cfTNA at <0.5% MAF were not detected in the matched tissue sample. These variants were also not detected in the cfTNA sample using another NGS assay or ddPCR assay suggesting that they were reported as false positives. The cfTNA input was below 20ng for these samples, demonstrating that the LoD and cut-off for reporting positive mutation results may vary based on cfTNA input.

Table 5. Variants Detected in Plasma Samples from Cancer Patients

Sample	Cancer type	cfTNA input (ng)	Oncomine PanCancer Cell Free Assay		Second assay A, B, or C for plasma and/or FFPE
			Variant	MAF (mut mol count)	
1	Melanoma	8	BRAF V600K	11.0% (173)	Confirmed in FFPE by assay A.
			PIK3CA E542K	0.18% (4)	Not detected in FFPE by assay A and B. Not detected in plasma by assay B.
2	Melanoma	10	NRAS Q61R	1.5% (24)	Confirmed in FFPE by assay A.
3	Gastric	10	TP53 R280G	4.2% (48)	Confirmed in FFPE by assay A.
4	Breast	8	BRAF V600E	0.18% (3)	Not detected in FFPE by assay A and C. Not detected in plasma by assay C.
5	CRC	28	No variants detected		No variants detected in plasma by assay B.
6	Breast	27	No variants detected		No variants detected in plasma by assay B.
7	Ovarian	17	PIK3CA E542K	0.34% (12)	Not detected in FFPE by assay A and B. Not detected in plasma by assay B.

Assays: A, OCAv3; B, Oncomine Breast cfDNA; C, ddPCR

Conclusions

- Oncomine Pan-Cancer Cell-Free Assay is able to detect multiple targets of interest in cfTNA including SNVs and indels, CNVs and fusions.
- SNVs and indels at 0.1% allele frequency can be successfully detected with 20 ng of input cfTNA.
- Additional validation is ongoing.

